



MANATEE COUNTY FLORIDA

September 5, 2014

TO: All Proposers

SUBJECT:

Request For Proposal (RFP) #14-2380JE
Discrete Analyzer Unit for Central Laboratory

ADDENDUM #1

The following items are issued to add to, modify and clarify the Request For Proposal document. Proposals are to be submitted on the specified time and date due, in conformance with the additions and revision listed herein.

A. Changes to Request For Proposal Structure

1. Section B.03 – 3.02 states the following: *"The instrument must have a 100% optical glass cuvette with a 10 mm path length."*

This requirement is not required.

2. Section B.03 – 3.17 states the following: *"The instrument must have the ability to add samples on the fly (ad-hoc) (i.e. must be able to add samples after sample analysis has begun without stopping or aborting run)."*

This statement is moved to Section B.02 – 2.5 due to it being a component of the System Overview and not a Technical Requirement.

No additional questions will be considered after the issuance of this Addendum.

Proposals are to be prepared as instructed in this Request For Proposals and shall be received at Manatee County Purchasing Office, Suite 803, 1112 Manatee Avenue West, Bradenton, Florida, FL 34205 until **4:00 P.M., September 24, 2014.**

Cordially,

Jacob Erickson
Contracts Negotiator

FINANCIAL MANAGEMENT – PURCHASING DIVISION
1112 Manatee Avenue West, Suite 803, Bradenton, FL 34205
PHONE: 941.749.3014 * FAX: 941.749.3034
www.myanatee.org